

**Table 4. Test characteristics of 64 slice CTA to detect ischemic disease in HF patients. (Adapted from Ghostine et al)**

<b>Classification</b>	<b>n</b>	<b>Sensitivity</b>	<b>Specificity</b>	<b>Accuracy</b>	<b>PPV</b>	<b>NPV</b>
<b>CAD present</b>	93	98 (88-100)	92 (81-98)	95 (88-98)	91 (79-98)	98 (89-100)
<b>ICMP present</b>	93	90 (74-98)	97 (89-100)	95 (88-98)	93 (78-99)	95 (87-99)